



FIG. 7.15 Test mandrel for Nb-Ti superconductor. (a) Exploded view of grooved fiberglass-epoxy thin-walled mandrel and copper current rings; (b) Sample wound on the mandrel and held in place, with the ends soldered to copper current rings; voltage leads are noninductively wound with the sample.